

Глюкометър Спецификации

Test Time	4 seconds	
Calibration Method	No coding	
Technology	Biosensor/Electrochemical, Glucose oxidase (GOD)	
Sample Type	Fresh whole blood	
Sample Size	0.4 µL	
Test Sites	Fingertip	
Hematocrit Range	25 - 60%	
Glucose Test Range	10 - 600 mg/dL (0.6 - 33.3 mmol/L)	
Memory Storage	500 results with date and time	
Test Averaging	7, 14, 30, 60, 90-day averages	
Data transfer	USB/Bluetooth	
Control Solution	3 levels	
Audio Feature	Optional beep for sample detection, error messages	
Automatic Shutoff	2 minutes after last action	
Batteries	One (1) CR 2032 3.0 V coin cell battery	
Operating Conditions	41 – 113 °F (5 – 45 °C) and 10 – 90% relative humidity	
Strip Shelf Life	24 months	
Open Vial Expiration	12 months	
Diabetes Management Software	Available	
Other Functions	Insufficient sample volume detection 2 nd sample application allowed	Meal markers – before & after meals Five (5) customizable test reminders